

INFORMATION BULLETIN

IFR 6000 XPDR/DME/TCAS/ADS-B/TIS/UAT TEST SET INFORMATION BULLETIN NO. 28 SOFTWARE RELEASE (V4.11.00)

I. PURPOSE:

The purpose of this information bulletin is to notify IFR 6000 customers that software Version 4.11.00 is now available.

II. DESCRIPTION:

Version 4.11.00 provides the following changes to the IFR 6000:

GENERAL changes:

Fixed calibration recovery from backup to help prevent Calibration default errors from appearing.

NOTE: This software version can only be used in IFR 6000 units that have MOD 3 marked on back of unit or a confirmed software version of 4.04.00 installed in the unit. This version is NOT backwards compatible for units with version 3.XX or less installed.

Mod 3 for the IFR 6000 is being used to indicate that the unit has an updated PC board installed and requires the 4.XX software version. The board was updated to incorporate the UAT function onto the one board instead of being on two.

IFR 6000 units with the Mod 3 do not have any additional features or capability when compared to the older IFR 6000 units with version 3.XX so updating an older unit is not necessary.

NOTE: Updating the software in the IFR 6000 with version 4.11.00 will not affect the calibration of the unit.

III. CONTACT INFORMATION:

Contact Cobham AvComm for more information.

Cobham AvComm
Customer Service Dept.
10200 W. York
Wichita, KS 67215

Telephone: (800) 835-2350
FAX: (316) 524-2623
Email: americas.service@aeroflex.com